



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: June 15, 2005 INITIALS: DEC OWNER: WSP
DISTRICT: 3 ADDRESS: High 12
LOCATION: Nashua #47 COUNTY: Yakima
MAKE: River Lake TYPE: truck scale ACCURACY CLASS: M4
MODEL: LQ 310 ID. NO.: 98438448 CAPACITY: 100,000 lbs SECTIONS: 2
MINIMUM GRADUATIONS: 20 lb REG. ELEMENT: electronic PLATFORM: 10' X 14'

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Sect 2</u>	<u>4000</u>	<u>4000</u>	<u>0</u>				
<u>7</u>	<u>8000</u>	<u>8000</u>	<u>0</u>	<u>Sect 1</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>
	<u>12000</u>	<u>12000</u>	<u>0</u>				
	<u>16000</u>	<u>16000</u>	<u>0</u>	<u>Center</u>	<u>20,000</u>	<u>20,000</u>	<u>0</u>
<u>Sect 2</u>	<u>20000</u>	<u>20000</u>	<u>0</u>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. L 2205-7 and 8

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Comer Date: June 16, 2005

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before
me on June 16, 2005 by

Michael J. Evans
(Signature of notary public)

Michael J. EVANS
Printed Name of Notary
My appointment expires 12-27-07

